	BID EVALUATION SHEET	Г	
Packag	e No/Tender Number :		
lame o	f The Equipment & Quantity : Oxygen Cylinders Qty-10, Estimated Cost. Rs. 75000 /- Each		
Part 1			
KNOCK	DOWN CRITERIA (COMMECIAL EVALUATION)		
All Eval	uation Parametrs defined below are mandatory for Complaince		
Sr No.	Evaluation Parameters	M/s Medlink Traders	M/s Total Technologies
1	Legal registered entity	Yes	Yes
2	Active Tax Payer (NTN & GST) as per clause 3 ITB	Yes	Yes
3	One person One Bid	Yes	Yes
4	Manufacturer/Sole Distributor	Yes	Yes
5	Past performance	Yes	Yes
6	Balance Sheet, Financial Statement & Annual Turnover	Yes	Yes
7	Complete item with accessories and optional	Yes	Yes
8	Certificate from the Manufacturer regarding after Sales Services and Installation	Yes	Yes
9	Technical & Engineering Capability of firm	Yes	Yes
10	Testing/Calibration equipment	Yes	Yes
11	Affidavit from Bidder (as per Supplier Declaration Template)	Yes	Yes
12	Bid Validity	Yes	Yes
13	Delivery Period	Yes	Yes
14	Compliance of Warranty as per tender	Yes	Yes
	REMARKS	Eligible for further evaluation of PART-II	Eligible for further evaluation of PART-II
Part II			
Knock [	Down Criteria (Product Evaluation)		
All Eval	uation Parametrs defined below are mandatory for Compliance		
Sr No	Specification Compliance / Evaluation Parameters		
27A	Oxygen Cylinders Larg	ge Size	

	Company Name	M/s Medlink Traders	M/s Total Technologies
	Brand	Oxytech	Worthington
	Model	0XY-0240	240 cft
	Country of Manufacturer	Oxytech	Austria
	Country of origin of product / Model Number	UK	Austria
	Product Local Market Business History as mentioned in criteria	Yes	Yes
	Compliance with defined quality standards (FDA 510k, MDD, Jp MHLW as mentioned in the specifications)	CE	CE
Sr#	Specifications/Sample Evaluation:		
1	240cft Medical cylinders	240CFT	240CFT
2	Nominal capacity 6800 ltrs/240cft	6800 Ltrs	6800 Ltrs
3	Nominal Pressure 137 Bars	137 Bar	137 Bar
4	Water capacity 47 ltrs	47 Ltrs	47 Ltrs
5	Bull nose valve	Yes	Yes
6	A handle wheel with transport safety cap shell be supplied with cylinders	Yes	Yes
7	Duly tested by renouned / authorized company	Yes	N0
8	BS specification & color code standard	Yes	Yes
9	Country of origin: USA/EU/Japan	UK	Austria
	TECHNICAL ELIGIBILITY OF THE PRODUCT	Eligible	Not-Eligible
	TECHNICAL ELIGIBILITY OF THE FIRM	Eligible	Eligible
	Bid Status	Responsive	Non-Responsive