

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR
UPGRADATION OF THQ HOSPITAL BUREWALA**

BID EVALUATION SHEET

Package No/Tender Number :

Name of The Equipment & Quantity : OXYGEN CYLINDERS Qty-10 each, Estimated Cost. Rs. 75000 /- Each

Part 1

KNOCK DOWN CRITERIA (COMMERCIAL EVALUATION)

All Evaluation Parametrs defined below are mandatory for Complainece

Sr No.	Evaluation Parameters	M/s Pakistan Healthcare & Medical Supplies	M/s Noorani Surgical	M/s Everest Technologies
1	Legal registered entity	Yes	Yes	Yes
2	Active Tax Payer (NTN & GST) as per clause 3 ITB	Yes	Yes	Yes
3	One person One Bid	Yes	Yes	Yes
4	Manufacturer/Sole Distributor	?	Yes	?
5	Past performance	?	Yes	?
6	Balance Sheet, Financial Statement & Annual Turnover	?	Yes	?
7	Complete item with accessories and optional	Yes	Yes	Yes
8	Certificate from the Manufacturer regarding after Sales Services and Installation	?	Yes	?
9	Technical & Engineering Capability of firm	?	Yes	?
10	Testing/Calibration equipment	?	Yes	?
11	Affidavit from Bidder (as per Supplier Declaration Template)	Yes	Yes	Yes
12	Bid Validity	Yes	Yes	Yes
13	Delivery Period	Yes	Yes	Yes
14	Compliance of Warranty as per tender	Yes	Yes	Yes
	REMARKS	Not-Eligible for further evaluation of PART-II	Eligible for further evaluation of PART-II	Not-Eligible for further evaluation of PART-II

Part II

Knock Down Criteria (Product Evaluation)				
All Evaluation Parametrs defined below are mandatory for Compliance				
Sr No	Specification Compliance / Evaluation Parameters			
2	OXYGEN CYLINDER - 240CFT			
	Company Name	M/s Pakistan Healthcare & Medical Supplies	M/s Noorani Surgical	M/s Everest Technologies
	Brand	Euro	Euro	Euro
	Model	240cft	240cft	240cft
	Country of Manufacturer	Germany	Germany	Germany
	Country of origin of product / Model Number	Germany	Germany	Germany
	Product Local Market Business History as mentioned in criteria	Yes	Yes	Yes
	Compliance with defined quality standards (FDA 510k, MDD, Jp MHLW as mentioned in the specifications)	Yes	Yes	Yes
Sr#	Specifications/Sample Evaluation:			
1	<u>TECHNICAL SPECIFICATIONS NOTE:</u>			
2	240cft Medical cylinders	Yes	Yes	Yes
3	Nominal capacity 6800 ltrs/240cft	Yes	Yes	Yes
4	Nominal Pressure 137 Bars	Yes	Yes	Yes
5	Water capacity 47 ltrs	Yes	Yes	Yes
6	Bull nose valve	Yes	Yes	Yes
7	A handle wheel with transport safety cap shell be supplied with cylinders	Yes	Yes	Yes
8	Duly tested by renouened / authorized company	Yes	Yes	Yes
9	BS specification & color code standard	Yes	Yes	Yes
10	Country of origin: USA/EU/Japan	EU	EU	EU
	TECHNICAL ELIGIBILITY OF THE PRODUCT	Eligible	Eligible	Eligible
	TECHNICAL ELIGIBILITY OF THE FIRM	Not-Eligible	Eligible	Not-Eligible
	Bid Status	Non-Responsive	Responsive	Non-Responsive