

**TENDER FOR THE PURCHASE OF MACHINERY, EQUIPMENT, FURNITURE & FIXTURE FOR  
UPGRADATION OF THQ HOSPITAL BUREWALA**

**BID EVALUATION SHEET**

Package No/Tender Number :

Name of The Equipment & Quantity : Patients Beds (Cardiac) + Over Bed Table + Bed Side Cabinet Qty-10 each, Estimated Cost. Rs. 169,500 /- Each Package

Part 1

**KNOCK DOWN CRITERIA ( COMMERCIAL EVALUATION)**

All Evaluation Parametrs defined below are mandatory for Complaine

Sr No.	Evaluation Parameters	M/s Pakistan Healthcare & Medical Supplies	M/s Noorani Surgical	M/s Everest Technologies
1	Legal registered entity	Yes	Yes	Yes
2	Active Tax Payer (NTN & GST) as per clause 3 ITB	Yes	Yes	Yes
3	One person One Bid	Yes	Yes	Yes
4	Manufacturer/Sole Distributor	?	?	?
5	Past performance	?	?	?
6	Balance Sheet, Financial Statement & Annual Turnover	?	Yes	?
7	Complete item with accessories and optional	Yes	Yes	Yes
8	Certificate from the Manufacturer regarding after Sales Services and Installation	?	?	?
9	Technical & Engineering Capability of firm	?	Yes	?
10	Testing/Calibration equipment	?	Yes	?
11	Affidavit from Bidder (as per Supplier Declaration Template)	Yes	Yes	Yes
12	Bid Validity	Yes	Yes	Yes
13	Delivery Period	Yes	Yes	Yes
14	Compliance of Warranty as per tender	Yes	Yes	Yes
	<b>REMARKS</b>	<b>Not-Eligible for further evaluation of PART-II</b>	<b>Not-Eligible for further evaluation of PART-II</b>	<b>Not-Eligible for further evaluation of PART-II</b>

Part II				
Knock Down Criteria (Product Evaluation)				
All Evaluation Parametrs defined below are mandatory for Compliance				
Sr No	Specification Compliance / Evaluation Parameters			
1	Patients Bed Cardiac			
	Company Name	M/s Pakistan Healthcare & Medical Supplies	M/s Noorani Surgical	M/s Everest Technologies
	Brand	Foshan Dongfang	Foshan Dongfang	Foshan Dongfang
	Model	?	FS3238WGZF4	?
	Country of Manufacturer	China	China	China
	Country of origin of product / Model Number	China	China	China
	Product Local Market Business History as mentioned in criteria	?	?	?
	Compliance with defined quality standards (FDA 510k, MDD, Jp MHLW as mentioned in the specifications)	?	?	?
Sr#	Specifications/Sample Evaluation:			
1	<b><u>TECHNICAL SPECIFICATIONS NOTE:</u></b>	Data sheet not provided	Data sheet not provided	Data sheet not provided
2	Electrically Operated Motorized Hospital bed suitable for use in Intensive Care Unit			
3	Height, head raise, knee bridge and tilt adjustment with hand held control and nurse control panel			
4	Size: L2210 x W1000 x H450/700mm			
5	Trendelenburg / anti Trendelenburg position			
6	1. Head Raise part : 24¾" x 33"			
7	2. Fixed Part : 7" x 33"			
8	3. Knee Bridge Part : 11½ x 33"			
9	4. Leg Rest part: 24½ "x 33" Bedding area made of			
10	ABS mesh.			
11	Radiolucent backrest			
12	Castors with bumpers			

13	Central Locking system			
14	Mattress with internal hinge and standard nylon proof cover			
15	Safe working load 350 KG			
16	<b>Finish:</b> All metals part to be chemically degreased / de-rusted through phosphate treatment and Covered with Antimicrobial polyester powder coating.			
17	<b>Welding:</b> All metal parts should be welded.			
18	<b>Side Rails:</b> Foldable side guards with smooth rounded plastic top and steel bars.			
19	Fixed to the bed on both sides covering 70% of platform length when engaged.			
20	Steel locking mechanism for strength.			
21	Minimum height when in use 30 cm from mattress platform.			
22	collapsible side rails			
23	Four swivel castors (all with individual brakes).			
24	Provision for I.V. Rod.			
25	<b>Over Bed Table:</b>			
26	Best Quality Subject to approval of sample			
27	<b>Bed Side Cabinet:</b>			
28	Best Quality Subject to approval of sample			
	<b>TECHNICAL ELIGIBILITY OF THE PRODUCT</b>	<b>Not-Eligible</b>	<b>Not-Eligible</b>	<b>Not-Eligible</b>
	<b>TECHNICAL ELIGIBILITY OF THE FIRM</b>	<b>Not-Eligible</b>	<b>Not-Eligible</b>	<b>Not-Eligible</b>
	<b>Bid Status</b>	<b>Non-Responsive</b>	<b>Non-Responsive</b>	<b>Non-Responsive</b>